

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060719\
 Data File : VX010133.D
 Acq On : 08 Jun 2019 00:51
 Operator : JC/SP
 Sample : K3257-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20190606

Quant Time: Jun 08 04:40:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	307964	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	469426	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	422746	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	200018	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.06	65	138201	45.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.94%	
35) Dibromofluoromethane	5.50	113	145175	50.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.28%	
50) Toluene-d8	8.71	98	553695	51.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.04%	
62) 4-Bromofluorobenzene	11.13	95	195522	48.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.48%	

Target Compounds

						Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	37212	13.740	ug/l	92
12) 1,1-Dichloroethene	2.37	96	21199	7.664	ug/l #	79
16) Acetone	2.44	43	3576	2.529	ug/l #	70
24) 1,1-Dichloroethane	3.70	63	5848	1.206	ug/l #	86
27) cis-1,2-Dichloroethene	4.60	96	9701	2.849	ug/l	71
30) Chloroform	5.21	83	2948	0.586	ug/l #	82
32) 1,1,1-Trichloroethane	5.50	97	2549	0.539	ug/l #	48
38) Carbon Tetrachloride	5.79	117	1882	0.434	ug/l	84
44) Trichloroethene	7.21	130	2548853	714.031	ug/l	91
49) 1,4-Dioxane	7.74	88	2579	29.031	ug/l #	80
55) 1,1,2-Trichloroethane	9.21	97	2274	0.713	ug/l	93
64) Tetrachloroethene	9.34	164	10292	3.278	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060719\
 Data File : VX010133.D
 Acq On : 08 Jun 2019 00:51
 Operator : JC/SP
 Sample : K3257-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20190606

Quant Time: Jun 08 04:40:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

