

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX060719\
 Data File : VX010135.D
 Acq On : 08 Jun 2019 01:37
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 04:44:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	50.000	53.037	-6.1	86	0.00
3 P	Chloromethane	50.000	51.864	-3.7	91	0.00
4 C	Vinyl Chloride	50.000	55.274	-10.5#	97	0.00
5 T	Bromomethane	50.000	53.829	-7.7	96	0.00
6 T	Chloroethane	50.000	57.763	-15.5	92	0.00
7 T	Trichlorofluoromethane	50.000	54.794	-9.6	98	0.00
8 T	Diethyl Ether	50.000	53.533	-7.1	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.044	3.9	88	0.00
10 T	Methyl Iodide	50.000	52.288	-4.6	90	0.00
11 T	Tert butyl alcohol	250.000	283.706	-13.5	101	0.00
12 CM	1,1-Dichloroethene	50.000	50.216	-0.4#	91	0.00
13 T	Acrolein	250.000	248.828	0.5	80	0.00
14 T	Allyl chloride	50.000	47.599	4.8	87	0.00
15 T	Acrylonitrile	250.000	261.240	-4.5	93	0.00
16 T	Acetone	250.000	219.627	12.1	80	0.00
17 T	Carbon Disulfide	50.000	47.513	5.0	88	0.00
18 T	Methyl Acetate	50.000	50.596	-1.2	94	0.00
19 T	Methyl tert-butyl Ether	50.000	49.330	1.3	90	0.00
20 T	Methylene Chloride	50.000	49.284	1.4	91	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.883	4.2	89	0.00
22 T	Diisopropyl ether	50.000	51.115	-2.2	91	0.00
23 T	Vinyl Acetate	250.000	256.266	-2.5	90	0.00
24 P	1,1-Dichloroethane	50.000	48.968	2.1	90	0.00
25 T	2-Butanone	250.000	255.017	-2.0	91	0.00
26 T	2,2-Dichloropropane	50.000	35.992	28.0#	66	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.484	-1.0	91	0.00
28 T	Bromochloromethane	50.000	47.440	5.1	86	0.00
29 T	Tetrahydrofuran	250.000	264.487	-5.8	96	0.00
30 C	Chloroform	50.000	50.351	-0.7#	91	0.00
31 T	Cyclohexane	50.000	50.083	-0.2	89	0.00
32 T	1,1,1-Trichloroethane	50.000	48.542	2.9	90	0.00
33 S	1,2-Dichloroethane-d4	50.000	52.005	-4.0	94	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	90	0.00
35 S	Dibromofluoromethane	50.000	52.718	-5.4	96	0.00
36 T	1,1-Dichloropropene	50.000	47.884	4.2	90	0.00
37 T	Ethyl Acetate	50.000	49.384	1.2	91	0.00
38 T	Carbon Tetrachloride	50.000	48.261	3.5	90	0.00
39 T	Methylcyclohexane	50.000	46.046	7.9	87	0.00
40 TM	Benzene	50.000	49.849	0.3	92	0.00
41 T	Methacrylonitrile	50.000	49.714	0.6	91	0.00
42 TM	1,2-Dichloroethane	50.000	48.289	3.4	89	0.00
43 T	Isopropyl Acetate	50.000	49.514	1.0	92	0.00
44 TM	Trichloroethene	50.000	51.313	-2.6	96	0.00
45 C	1,2-Dichloropropane	50.000	50.168	-0.3#	92	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX060719\
 Data File : VX010135.D
 Acq On : 08 Jun 2019 01:37
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 04:44:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 T	Dibromomethane	50.000	47.246	5.5	90	0.00
47 T	Bromodichloromethane	50.000	49.035	1.9	90	0.00
48 T	Methyl methacrylate	50.000	50.524	-1.0	90	0.00
49 T	1,4-Dioxane	1000.000	1009.071	-0.9	94	0.00
50 S	Toluene-d8	50.000	53.638	-7.3	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	263.960	-5.6	94	0.00
52 CM	Toluene	50.000	49.157	1.7#	92	0.00
53 T	t-1,3-Dichloropropene	50.000	46.904	6.2	86	0.00
54 T	cis-1,3-Dichloropropene	50.000	46.697	6.6	86	0.00
55 T	1,1,2-Trichloroethane	50.000	51.352	-2.7	94	0.00
56 T	Ethyl methacrylate	50.000	51.390	-2.8	92	0.00
57 T	1,3-Dichloropropane	50.000	49.557	0.9	92	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	244.613	2.2	87	0.00
59 T	2-Hexanone	250.000	257.289	-2.9	92	0.00
60 T	Dibromochloromethane	50.000	50.302	-0.6	92	0.00
61 T	1,2-Dibromoethane	50.000	49.989	0.0	93	0.00
62 S	4-Bromofluorobenzene	50.000	53.412	-6.8	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	90	0.00
64 T	Tetrachloroethene	50.000	50.659	-1.3	98	0.00
65 PM	Chlorobenzene	50.000	48.879	2.2	91	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.927	0.1	93	0.00
67 C	Ethyl Benzene	50.000	49.081	1.8#	92	0.00
68 T	m/p-Xylenes	100.000	98.907	1.1	92	0.00
69 T	o-Xylene	50.000	48.767	2.5	92	0.00
70 T	Styrene	50.000	49.691	0.6	91	0.00
71 P	Bromoform	50.000	49.341	1.3	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	90	0.00
73 T	Isopropylbenzene	50.000	49.809	0.4	92	0.00
74 T	N-amyl acetate	50.000	50.501	-1.0	93	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.393	1.2	93	0.00
76 T	1,2,3-Trichloropropane	50.000	48.551	2.9	91	0.00
77 T	Bromobenzene	50.000	49.238	1.5	92	0.00
78 T	n-propylbenzene	50.000	49.596	0.8	91	0.00
79 T	2-Chlorotoluene	50.000	49.306	1.4	92	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.475	1.0	92	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.722	8.6	84	0.00
82 T	4-Chlorotoluene	50.000	49.761	0.5	91	0.00
83 T	tert-Butylbenzene	50.000	49.232	1.5	92	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.428	1.1	92	0.00
85 T	sec-Butylbenzene	50.000	49.269	1.5	92	0.00
86 T	p-Isopropyltoluene	50.000	49.109	1.8	91	0.00
87 T	1,3-Dichlorobenzene	50.000	48.886	2.2	91	0.00
88 T	1,4-Dichlorobenzene	50.000	48.097	3.8	92	0.00
89 T	n-Butylbenzene	50.000	48.605	2.8	89	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060719\
Data File : VX010135.D
Acq On : 08 Jun 2019 01:37
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 08 04:44:44 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	50.000	49.358	1.3	92	0.00
91 T	1,2-Dichlorobenzene	50.000	49.513	1.0	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.337	5.3	91	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.339	5.3	91	0.00
94 T	Hexachlorobutadiene	50.000	45.658	8.7	87	0.00
95 T	Naphthalene	50.000	49.035	1.9	92	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.272	3.5	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6