

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060719\
 Data File : VX010119.D
 Acq On : 07 Jun 2019 19:24
 Operator : JC/SP
 Sample : K3221-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW313S-20190605

Quant Time: Jun 08 03:49:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	335687	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	510002	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	459434	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	216882	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	156371	47.72	ug/l	0.00
Spiked Amount			Recovery	=	95.44%	
35) Dibromofluoromethane	5.50	113	159439	50.68	ug/l	0.00
Spiked Amount			Recovery	=	101.36%	
50) Toluene-d8	8.71	98	600694	50.95	ug/l	0.00
Spiked Amount			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	11.13	95	214237	48.66	ug/l	0.00
Spiked Amount			Recovery	=	97.32%	


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060719\
Data File : VX010119.D
Acq On : 07 Jun 2019 19:24
Operator : JC/SP
Sample : K3221-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW313S-20190605

Quant Time: Jun 08 03:49:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX010119.D

Acq: 07 Jun 2019 19:24

Instrument :

MSVOA_X

ClientSampleId :


BPS1-TT-MW313S-20190605

Tot Ion:168 Resp: 335687

Ion Ratio Lower Upper




168 100

99 45.5 43.0 64.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 47.722 ug/l

RT: 6.06 min Scan# 805

Delta R.T. 0.00 min

Lab File: VX010119.D


Acq: 07 Jun 2019 19:24

Tgt Ion: 65 Resp: 156371

Ion Ratio Lower Upper



65 100


67 53.9 0.0 108.4



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX010119.D

Acq: 07 Jun 2019 19:24

Instrument :

MSVOA_X

ClientSampleId :

BPS1-TT-MW313S-20190605


Tot Ion:114 Resp: 510002

Ion Ratio Lower Upper

114 100

63 17.1 0.0 47.4




88 14.7 0.0 33.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 50.683 ug/l

RT: 5.50 min Scan# 713

Delta R.T. 0.00 min

Lab File: VX010119.D

Acq: 07 Jun 2019 19:24


Tgt Ion:113 Resp: 159439

Ion Ratio Lower Upper

113 100

111 101.4 82.2 123.2



192 22.8 17.1 25.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 50.949 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX010119.D

Acq: 07 Jun 2019 19:24

Instrument :

MSVOA_X

ClientSampleId :


BPS1-TT-MW313S-20190605

Tot Ion: 98 Resp: 600694

Ion Ratio Lower Upper




98 100

100 65.0 51.5 77.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 48.656 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX010119.D

Acq: 07 Jun 2019 19:24


Tgt Ion: 95 Resp: 214237

Ion Ratio Lower Upper

95 100

174 90.0 0.0 160.4



176 87.3 0.0 154.6




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX010119.D
Acq: 07 Jun 2019 19:24

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW313S-20190605

Tot Ion:117 Resp: 459434
Ion Ratio Lower Upper
117 100
82 51.2 49.2 73.8
119 32.5 25.2 37.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX010119.D
Acq: 07 Jun 2019 19:24

Tgt Ion:152 Resp: 216882
Ion Ratio Lower Upper
152 100
115 54.7 41.4 124.2
150 156.8 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

