

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060721\
 Data File : VX022534.D
 Acq On : 07 Jun 2021 15:21
 Operator : JC/MD
 Sample : VX0607MBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0607MBL01

Quant Time: Jun 07 17:04:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	67661	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	141524	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46403	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56144	46.676	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.360%
35) Dibromofluoromethane	5.391	113	41281	46.621	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.240%
50) Toluene-d8	8.653	98	168592	47.718	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.440%
62) 4-Bromofluorobenzene	11.085	95	57508	43.791	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060721\
Data File : VX022534.D
Acq On : 07 Jun 2021 15:21
Operator : JC/MD
Sample : VX0607MBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0607MBL01

Quant Time: Jun 07 17:04:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

