

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060721\
 Data File : VX022550.D
 Acq On : 07 Jun 2021 21:35
 Operator : JC/MD
 Sample : M2615-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1245

Quant Time: Jun 08 04:23:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59812	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	123025	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	109032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44072	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51552	48.483	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.960%
35) Dibromofluoromethane	5.403	113	35615	46.270	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.540%
50) Toluene-d8	8.653	98	149987	48.836	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.085	95	53145	46.554	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.100%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	13807	28.898	ug/l	96
20) Methylene Chloride	2.794	84	1351	1.354	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060721\
Data File : VX022550.D
Acq On : 07 Jun 2021 21:35
Operator : JC/MD
Sample : M2615-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1245

Quant Time: Jun 08 04:23:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

