



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060721\
 Data File : VX022524.D
 Acq On : 07 Jun 2021 10:13
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Title : SW846 8260
 Last Update : Mon Jun 07 14:27:50 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.0	5678	PASS
75	95	30	60	53.6	15237	PASS
95	95	100	100	100.0	28405	PASS
96	95	5	9	6.6	1871	PASS
173	174	0.00	2	1.0	178	PASS
174	95	50	100	65.0	18461	PASS
175	174	5	9	6.9	1265	PASS
176	174	95	101	97.9	18078	PASS
177	176	5	9	6.3	1131	PASS