

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
 Data File : VX029277.D
 Acq On : 07 Jun 2022 13:48
 Operator : JC/MD
 Sample : VX0607WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0607WBL01

Quant Time: Jun 08 01:23:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	399426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	664518	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	578023	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	249149	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	236969	52.612	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.220%
35) Dibromofluoromethane	5.385	113	199762	45.258	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.520%
50) Toluene-d8	8.653	98	752371	46.606	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.220%
62) 4-Bromofluorobenzene	11.079	95	266009	42.628	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
Data File : VX029277.D
Acq On : 07 Jun 2022 13:48
Operator : JC/MD
Sample : VX0607WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0607WBL01

Quant Time: Jun 08 01:23:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

