

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
 Data File : VX029282.D
 Acq On : 07 Jun 2022 16:41
 Operator : JC/MD
 Sample : PB145355TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145355TB

Quant Time: Jun 08 01:25:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	395104	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	655040	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	583314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	245819	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	239512	53.758	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.520%
35) Dibromofluoromethane	5.391	113	200359	46.050	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.100%
50) Toluene-d8	8.653	98	759250	47.712	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.420%
62) 4-Bromofluorobenzene	11.085	95	266505	43.325	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.660%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	40382	23.017	ug/l	91
20) Methylene Chloride	2.794	84	15812	3.866	ug/l	95
43) Isopropyl Acetate	6.348	43	30393	3.581	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
Data File : VX029282.D
Acq On : 07 Jun 2022 16:41
Operator : JC/MD
Sample : PB145355TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145355TB

Quant Time: Jun 08 01:25:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

