

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
 Data File : VX029300.D
 Acq On : 08 Jun 2022 01:54
 Operator : JC/MD
 Sample : PB145355ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145355ZHE#04

Quant Time: Jun 08 03:19:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	414161	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	699010	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	614378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	255910	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	255177	40.130	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	80.260%
35) Dibromofluoromethane	5.391	113	212422	35.746	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	71.500%
50) Toluene-d8	8.653	98	802793	38.197	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	76.400%
62) 4-Bromofluorobenzene	11.085	95	283212	34.434	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	68.860%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	65044	19.414	ug/l	93
20) Methylene Chloride	2.794	84	27690	3.701	ug/l	96
43) Isopropyl Acetate	6.348	43	28493	1.872	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
Data File : VX029300.D
Acq On : 08 Jun 2022 01:54
Operator : JC/MD
Sample : PB145355ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145355ZHE#04

Quant Time: Jun 08 03:19:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

