

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
 Data File : VX029301.D
 Acq On : 08 Jun 2022 02:17
 Operator : JC/MD
 Sample : PB145355ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145355ZHE#05

Quant Time: Jun 08 03:20:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	398856	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	685117	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	602994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	260963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	251861	41.173	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	82.340%
35) Dibromofluoromethane	5.391	113	207515	35.628	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	71.260%
50) Toluene-d8	8.653	98	788562	38.281	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	76.560%
62) 4-Bromofluorobenzene	11.085	95	281173	34.879	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	69.760%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	66640	20.653	ug/l	96
20) Methylene Chloride	2.788	84	27016	3.749	ug/l	98
43) Isopropyl Acetate	6.342	43	29939	2.007	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
Data File : VX029301.D
Acq On : 08 Jun 2022 02:17
Operator : JC/MD
Sample : PB145355ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145355ZHE#05

Quant Time: Jun 08 03:20:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

