

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
 Data File : VX029305.D
 Acq On : 08 Jun 2022 03:50
 Operator : JC/MD
 Sample : PB145355ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145355ZHE#09

Quant Time: Jun 08 05:17:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	398869	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	670348	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	597068	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	248747	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	252612	41.300	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	82.600%
35) Dibromofluoromethane	5.391	113	205708	36.096	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	72.200%
50) Toluene-d8	8.653	98	777245	38.563	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	77.120%
62) 4-Bromofluorobenzene	11.079	95	275953	34.986	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	69.980%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	44074	13.659	ug/l	94
20) Methylene Chloride	2.788	84	15071	2.091	ug/l	90
43) Isopropyl Acetate	6.348	43	30723	2.104	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
Data File : VX029305.D
Acq On : 08 Jun 2022 03:50
Operator : JC/MD
Sample : PB145355ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145355ZHE#09

Quant Time: Jun 08 05:17:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

