

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
 Data File : VX029309.D
 Acq On : 08 Jun 2022 05:23
 Operator : JC/MD
 Sample : N3194-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1

Quant Time: Jun 08 09:19:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	387539	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	652103	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	585072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	247141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	244208	41.084	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	82.160%
35) Dibromofluoromethane	5.391	113	197426	35.612	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	71.220%
50) Toluene-d8	8.653	98	756127	38.564	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	77.120%
62) 4-Bromofluorobenzene	11.079	95	271922	35.439	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	70.880%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	43674	13.931	ug/l	94
20) Methylene Chloride	2.788	84	15208	2.172	ug/l	94
43) Isopropyl Acetate	6.348	43	29355	2.067	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
Data File : VX029309.D
Acq On : 08 Jun 2022 05:23
Operator : JC/MD
Sample : N3194-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1

Quant Time: Jun 08 09:19:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

