

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX060722\
 Data File : VX029292.D
 Acq On : 07 Jun 2022 22:02
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX060722

Quant Time: Jun 08 03:18:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	50.000	48.987	2.0	99	0.00
3 P	Chloromethane	50.000	45.979	8.0	99	0.00
4 C	Vinyl Chloride	50.000	48.342	3.3#	98	0.00
5 T	Bromomethane	50.000	54.237	-8.5	117	0.00
6 T	Chloroethane	50.000	48.495	3.0	94	0.00
7 T	Trichlorofluoromethane	50.000	48.549	2.9	98	0.00
8 T	Diethyl Ether	50.000	47.708	4.6	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.100	3.8	100	0.00
10 T	Methyl Iodide	50.000	40.939	18.1	83	0.00
11 T	Tert butyl alcohol	250.000	256.532	-2.6	100	0.00
12 CM	1,1-Dichloroethene	50.000	48.063	3.9#	98	0.00
13 T	Acrolein	250.000	232.675	6.9	101	0.00
14 T	Allyl chloride	50.000	48.114	3.8	98	0.00
15 T	Acrylonitrile	250.000	238.587	4.6	97	0.00
16 T	Acetone	250.000	230.845	7.7	98	0.00
17 T	Carbon Disulfide	50.000	48.712	2.6	99	0.00
18 T	Methyl Acetate	50.000	46.003	8.0	96	0.00
19 T	Methyl tert-butyl Ether	50.000	49.281	1.4	100	0.00
20 T	Methylene Chloride	50.000	45.652	8.7	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.163	3.7	100	0.00
22 T	Diisopropyl ether	50.000	49.022	2.0	97	0.00
23 T	Vinyl Acetate	250.000	254.889	-2.0	97	0.00
24 P	1,1-Dichloroethane	50.000	47.717	4.6	97	0.00
25 T	2-Butanone	250.000	239.793	4.1	96	0.00
26 T	2,2-Dichloropropane	50.000	46.490	7.0	94	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.186	5.6	99	0.00
28 T	Bromochloromethane	50.000	49.546	0.9	101	0.00
29 T	Tetrahydrofuran	250.000	243.412	2.6	96	0.00
30 C	Chloroform	50.000	47.690	4.6#	98	0.00
31 T	Cyclohexane	50.000	48.288	3.4	99	0.00
32 T	1,1,1-Trichloroethane	50.000	48.767	2.5	98	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.108	3.8	105	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	107	0.00
35 S	Dibromofluoromethane	50.000	45.621	8.8	105	0.00
36 T	1,1-Dichloropropene	50.000	48.055	3.9	97	0.00
37 T	Ethyl Acetate	50.000	49.238	1.5	96	0.00
38 T	Carbon Tetrachloride	50.000	48.891	2.2	99	0.00
39 T	Methylcyclohexane	50.000	48.689	2.6	98	0.00
40 TM	Benzene	50.000	48.712	2.6	99	0.00
41 T	Methacrylonitrile	50.000	51.822	-3.6	100	0.00
42 TM	1,2-Dichloroethane	50.000	47.585	4.8	96	0.00
43 T	Isopropyl Acetate	50.000	49.905	0.2	98	0.00
44 TM	Trichloroethene	50.000	47.994	4.0	100	0.00
45 C	1,2-Dichloropropane	50.000	48.272	3.5#	98	0.00
46 T	Dibromomethane	50.000	48.782	2.4	98	0.00
47 T	Bromodichloromethane	50.000	48.960	2.1	97	0.00
48 T	Methyl methacrylate	50.000	48.796	2.4	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
 Data File : VX029292.D
 Acq On : 07 Jun 2022 22:02
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX060722

Quant Time: Jun 08 03:18:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1005.089	-0.5	97	0.00
50 S	Toluene-d8	50.000	45.978	8.0	106	0.00
51 T	4-Methyl-2-Pentanone	250.000	251.641	-0.7	97	0.00
52 CM	Toluene	50.000	49.843	0.3#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	47.625	4.8	99	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.037	-2.1	97	0.00
55 T	1,1,2-Trichloroethane	50.000	48.922	2.2	97	0.00
56 T	Ethyl methacrylate	50.000	52.165	-4.3	97	0.00
57 T	1,3-Dichloropropane	50.000	48.284	3.4	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	255.798	-2.3	99	0.00
59 T	2-Hexanone	250.000	256.112	-2.4	97	0.00
60 T	Dibromochloromethane	50.000	50.951	-1.9	98	0.00
61 T	1,2-Dibromoethane	50.000	48.620	2.8	96	0.00
62 S	4-Bromofluorobenzene	50.000	47.314	5.4	107	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	106	0.00
64 T	Tetrachloroethene	50.000	49.310	1.4	99	0.00
65 PM	Chlorobenzene	50.000	48.696	2.6	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.926	-1.9	98	0.00
67 C	Ethyl Benzene	50.000	49.733	0.5#	97	0.00
68 T	m/p-Xylenes	100.000	101.061	-1.1	98	0.00
69 T	o-Xylene	50.000	51.287	-2.6	99	0.00
70 T	Styrene	50.000	51.193	-2.4	98	0.00
71 P	Bromoform	50.000	51.281	-2.6	96	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	106	0.00
73 T	Isopropylbenzene	50.000	49.271	1.5	97	0.00
74 T	N-amyl acetate	50.000	52.677	-5.4	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.313	5.4	96	0.00
76 T	1,2,3-Trichloropropane	50.000	48.843	2.3	96	0.00
77 T	Bromobenzene	50.000	48.678	2.6	98	0.00
78 T	n-propylbenzene	50.000	49.707	0.6	97	0.00
79 T	2-Chlorotoluene	50.000	49.079	1.8	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.630	0.7	97	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.473	7.1	94	0.00
82 T	4-Chlorotoluene	50.000	48.862	2.3	97	0.00
83 T	tert-Butylbenzene	50.000	49.729	0.5	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.154	-0.3	98	0.00
85 T	sec-Butylbenzene	50.000	49.920	0.2	97	0.00
86 T	p-Isopropyltoluene	50.000	50.825	-1.7	98	0.00
87 T	1,3-Dichlorobenzene	50.000	49.145	1.7	99	0.00
88 T	1,4-Dichlorobenzene	50.000	48.256	3.5	99	0.00
89 T	n-Butylbenzene	50.000	50.764	-1.5	97	0.00
90 T	Hexachloroethane	50.000	50.157	-0.3	99	0.00
91 T	1,2-Dichlorobenzene	50.000	48.604	2.8	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.508	1.0	96	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.491	-1.0	100	0.00
94 T	Hexachlorobutadiene	50.000	48.647	2.7	100	0.00
95 T	Naphthalene	50.000	52.047	-4.1	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
Data File : VX029292.D
Acq On : 07 Jun 2022 22:02
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX060722

Quant Time: Jun 08 03:18:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.894	0.2	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6