

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX060722\
 Data File : VX029311.D
 Acq On : 08 Jun 2022 06:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 09:23:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	50.000	50.968	-1.9	93	0.00
3 P	Chloromethane	50.000	50.140	-0.3	97	0.00
4 C	Vinyl Chloride	50.000	51.661	-3.3#	95	0.00
5 T	Bromomethane	50.000	52.214	-4.4	101	0.00
6 T	Chloroethane	50.000	50.871	-1.7	89	0.00
7 T	Trichlorofluoromethane	50.000	51.231	-2.5	94	0.00
8 T	Diethyl Ether	50.000	50.713	-1.4	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.074	-0.1	94	0.00
10 T	Methyl Iodide	50.000	47.316	5.4	87	0.00
11 T	Tert butyl alcohol	250.000	262.637	-5.1	92	0.00
12 CM	1,1-Dichloroethene	50.000	50.404	-0.8#	93	0.00
13 T	Acrolein	250.000	216.364	13.5	85	0.00
14 T	Allyl chloride	50.000	49.141	1.7	90	0.00
15 T	Acrylonitrile	250.000	248.323	0.7	91	0.00
16 T	Acetone	250.000	245.537	1.8	94	0.00
17 T	Carbon Disulfide	50.000	50.302	-0.6	93	0.00
18 T	Methyl Acetate	50.000	48.885	2.2	92	0.00
19 T	Methyl tert-butyl Ether	50.000	51.944	-3.9	95	0.00
20 T	Methylene Chloride	50.000	47.834	4.3	93	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.324	-0.6	94	0.00
22 T	Diisopropyl ether	50.000	53.617	-7.2	96	0.00
23 T	Vinyl Acetate	250.000	272.634	-9.1	94	0.00
24 P	1,1-Dichloroethane	50.000	51.166	-2.3	94	0.00
25 T	2-Butanone	250.000	250.391	-0.2	91	0.00
26 T	2,2-Dichloropropane	50.000	37.588	24.8	69	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.460	1.1	94	0.00
28 T	Bromochloromethane	50.000	51.562	-3.1	95	0.00
29 T	Tetrahydrofuran	250.000	257.150	-2.9	92	0.00
30 C	Chloroform	50.000	50.772	-1.5#	94	0.00
31 T	Cyclohexane	50.000	51.374	-2.7	95	0.00
32 T	1,1,1-Trichloroethane	50.000	51.456	-2.9	93	0.00
33 S	1,2-Dichloroethane-d4	50.000	52.462	-4.9	103	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	99	0.00
35 S	Dibromofluoromethane	50.000	47.237	5.5	100	0.00
36 T	1,1-Dichloropropene	50.000	49.925	0.2	93	0.00
37 T	Ethyl Acetate	50.000	50.890	-1.8	92	0.00
38 T	Carbon Tetrachloride	50.000	50.053	-0.1	94	0.00
39 T	Methylcyclohexane	50.000	49.816	0.4	93	0.00
40 TM	Benzene	50.000	49.789	0.4	94	0.00
41 T	Methacrylonitrile	50.000	52.912	-5.8	95	0.00
42 TM	1,2-Dichloroethane	50.000	50.131	-0.3	94	0.00
43 T	Isopropyl Acetate	50.000	51.750	-3.5	94	0.00
44 TM	Trichloroethene	50.000	48.333	3.3	93	0.00
45 C	1,2-Dichloropropane	50.000	50.618	-1.2#	95	0.00
46 T	Dibromomethane	50.000	49.630	0.7	93	0.00
47 T	Bromodichloromethane	50.000	51.302	-2.6	94	0.00
48 T	Methyl methacrylate	50.000	51.159	-2.3	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
 Data File : VX029311.D
 Acq On : 08 Jun 2022 06:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 09:23:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	996.530	0.3	89	0.00
50 S	Toluene-d8	50.000	47.052	5.9	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	260.772	-4.3	93	0.00
52 CM	Toluene	50.000	50.761	-1.5#	93	0.00
53 T	t-1,3-Dichloropropene	50.000	47.554	4.9	92	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.435	-0.9	89	0.00
55 T	1,1,2-Trichloroethane	50.000	50.394	-0.8	93	0.00
56 T	Ethyl methacrylate	50.000	53.903	-7.8	93	0.00
57 T	1,3-Dichloropropane	50.000	50.283	-0.6	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	265.202	-6.1	95	0.00
59 T	2-Hexanone	250.000	262.446	-5.0	92	0.00
60 T	Dibromochloromethane	50.000	51.772	-3.5	92	0.00
61 T	1,2-Dibromoethane	50.000	50.486	-1.0	92	0.00
62 S	4-Bromofluorobenzene	50.000	48.897	2.2	103	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	100	0.00
64 T	Tetrachloroethene	50.000	44.456	11.1	84	0.00
65 PM	Chlorobenzene	50.000	48.704	2.6	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.299	-2.6	93	0.00
67 C	Ethyl Benzene	50.000	50.748	-1.5#	93	0.00
68 T	m/p-Xylenes	100.000	102.349	-2.3	93	0.00
69 T	o-Xylene	50.000	51.630	-3.3	93	0.00
70 T	Styrene	50.000	52.008	-4.0	94	0.00
71 P	Bromoform	50.000	51.026	-2.1	90	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	101	0.00
73 T	Isopropylbenzene	50.000	49.408	1.2	93	0.00
74 T	N-ethyl acetate	50.000	53.922	-7.8	96	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.550	2.9	94	0.00
76 T	1,2,3-Trichloropropane	50.000	48.543	2.9	91	0.00
77 T	Bromobenzene	50.000	48.078	3.8	93	0.00
78 T	n-propylbenzene	50.000	50.263	-0.5	93	0.00
79 T	2-Chlorotoluene	50.000	49.585	0.8	94	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.668	0.7	93	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	42.735	14.5	82	0.00
82 T	4-Chlorotoluene	50.000	49.412	1.2	94	0.00
83 T	tert-Butylbenzene	50.000	50.536	-1.1	94	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.674	-1.3	94	0.00
85 T	sec-Butylbenzene	50.000	50.324	-0.6	93	0.00
86 T	p-Isopropyltoluene	50.000	50.180	-0.4	93	0.00
87 T	1,3-Dichlorobenzene	50.000	48.436	3.1	93	0.00
88 T	1,4-Dichlorobenzene	50.000	47.850	4.3	94	0.00
89 T	n-Butylbenzene	50.000	50.565	-1.1	92	0.00
90 T	Hexachloroethane	50.000	50.087	-0.2	94	0.00
91 T	1,2-Dichlorobenzene	50.000	48.087	3.8	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.025	4.0	89	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.282	3.4	91	0.00
94 T	Hexachlorobutadiene	50.000	44.676	10.6	87	0.00
95 T	Naphthalene	50.000	50.213	-0.4	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060722\
Data File : VX029311.D
Acq On : 08 Jun 2022 06:09
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 08 09:23:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.250	3.5	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6