

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
 Data File : VX036054.D
 Acq On : 07 Jun 2023 12:36
 Operator : JC/MD
 Sample : VX0607WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0607WBL01

Quant Time: Jun 08 07:22:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	40924	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	6.757	114	231968	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	204142	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	108561	31.926	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 106.433%	
60) 4-Bromofluorobenzene	11.079	95	104803	30.595	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 102.000%	
63) Toluene-d8	8.647	98	271752	30.112	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 100.367%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
Data File : VX036054.D
Acq On : 07 Jun 2023 12:36
Operator : JC/MD
Sample : VX0607WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0607WBL01

Quant Time: Jun 08 07:22:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

