

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
 Data File : VX036059.D
 Acq On : 07 Jun 2023 14:39
 Operator : JC/MD
 Sample : 03016-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-45(OW)

Quant Time: Jun 08 07:22:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	32649	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	6.757	114	177605	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	161862	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	93216	34.362	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	114.533%#
60) 4-Bromofluorobenzene	11.079	95	81707	30.083	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	100.267%
63) Toluene-d8	8.647	98	213064	29.776	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.267%
Target Compounds						
						Qvalue
15) Acetone	2.398	58	2051	5.915	ug/l	83
16) Carbon Disulfide	2.514	76	1506	0.424	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
Data File : VX036059.D
Acq On : 07 Jun 2023 14:39
Operator : JC/MD
Sample : 03016-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-45(OW)

Quant Time: Jun 08 07:22:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

