

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
 Data File : VX036061.D
 Acq On : 07 Jun 2023 15:26
 Operator : JC/MD
 Sample : 03016-08RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-45(OW)RE

Quant Time: Jun 08 07:22:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	30731	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	6.763	114	163421	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	151902	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	85085	33.322	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	111.067%#
60) 4-Bromofluorobenzene	11.079	95	77026	30.220	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	100.733%
63) Toluene-d8	8.647	98	194684	28.991	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.633%
Target Compounds						
15) Acetone	2.404	58	1964	6.017	ug/l	Qvalue # 57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
Data File : VX036061.D
Acq On : 07 Jun 2023 15:26
Operator : JC/MD
Sample : 03016-08RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-45(OW)RE

Quant Time: Jun 08 07:22:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

