

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
 Data File : VX036062.D
 Acq On : 07 Jun 2023 15:49
 Operator : JC/MD
 Sample : 03047-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-41

Quant Time: Jun 08 07:22:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/08/2023
 Supervised By : Mahesh Dadoda 06/08/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	21313	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	6.757	114	120433	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	104081	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	58801	33.204	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	110.667%#
60) 4-Bromofluorobenzene	11.079	95	51820	29.672	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	98.900%
63) Toluene-d8	8.647	98	138719	30.148	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.500%
Target Compounds						
3) Chloromethane	1.301	50	464m	0.347	ug/l	Qvalue
15) Acetone	2.404	58	852m	3.764	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
Data File : VX036062.D
Acq On : 07 Jun 2023 15:49
Operator : JC/MD
Sample : 03047-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-41

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/08/2023
Supervised By : Mahesh Dadoda 06/08/2023

Quant Time: Jun 08 07:22:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

