

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
 Data File : VX036063.D
 Acq On : 07 Jun 2023 16:12
 Operator : JC/MD
 Sample : 03047-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP-07

Quant Time: Jun 08 07:22:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	36646	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	192580	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	176169	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	94797	31.133	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.767%
60) 4-Bromofluorobenzene	11.079	95	84706	28.655	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	95.500%
63) Toluene-d8	8.647	98	229518	29.470	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.233%
Target Compounds						
					Qvalue	
3) Chloromethane	1.300	50	1556	0.677	ug/l #	88
15) Acetone	2.410	58	2465	6.333	ug/l	72
16) Carbon Disulfide	2.514	76	1391	0.349	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
Data File : VX036063.D
Acq On : 07 Jun 2023 16:12
Operator : JC/MD
Sample : 03047-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP-07

Quant Time: Jun 08 07:22:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

