

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
 Data File : VX036065.D
 Acq On : 07 Jun 2023 16:59
 Operator : JC/MD
 Sample : 03047-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-41

Quant Time: Jun 08 07:23:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	33472	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	177050	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	166230	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	90958	32.705	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	109.000%
60) 4-Bromofluorobenzene	11.079	95	80841	28.983	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	96.600%
63) Toluene-d8	8.647	98	214564	29.197	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.333%
Target Compounds						
					Qvalue	
3) Chloromethane	1.301	50	480	0.229	ug/l #	85
15) Acetone	2.404	58	862	2.425	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
Data File : VX036065.D
Acq On : 07 Jun 2023 16:59
Operator : JC/MD
Sample : 03047-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-41

Quant Time: Jun 08 07:23:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

