

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
 Data File : VX036066.D
 Acq On : 07 Jun 2023 17:22
 Operator : JC/MD
 Sample : 03069-11RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP08RE

Quant Time: Jun 08 07:23:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	34244	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	180053	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	161502	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	95463	33.551	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	111.833%#
60) 4-Bromofluorobenzene	11.079	95	81905	30.224	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	100.733%
63) Toluene-d8	8.647	98	213896	29.959	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.867%
Target Compounds						
15) Acetone	2.398	58	1267	3.484	ug/l	Qvalue 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
Data File : VX036066.D
Acq On : 07 Jun 2023 17:22
Operator : JC/MD
Sample : 03069-11RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP08RE

Quant Time: Jun 08 07:23:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

