

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
 Data File : VX036067.D
 Acq On : 07 Jun 2023 17:46
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Jun 08 07:23:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	99	0.00
2 M	Dichlorodifluoromethane	1.534	1.949	-27.1#	196#	0.00
3 M	Chloromethane	1.882	1.834	2.6	109	0.00
4 M	Vinyl Chloride	1.769	1.794	-1.4	122	0.00
5 M	Bromomethane	1.190	0.937	21.3	85	0.00
6 M	Chloroethane	1.115	1.225	-9.9	125	0.00
7 M	Trichlorofluoromethane	2.681	2.899	-8.1	150#	0.00
8 T	Diethyl Ether	1.115	1.085	2.7	102	0.00
9	1,1,2-Trichlorotrifluoroeth	1.443	1.737	-20.4	180#	0.00
10 M	1,1-Dichloroethene	1.504	1.585	-5.4	124	0.00
11	Methyl Iodide	1.817	1.113	38.7#	70	0.00
12	Methyl Acetate	4.249	4.754	-11.9	113	0.00
13 M	Acrolein	0.336	0.304	9.5	99	0.00
14 M	Acrylonitrile	1.065	1.142	-7.2	110	0.00
15 M	Acetone	0.319	0.319	0.0	99	0.00
16 M	Carbon Disulfide	3.264	3.172	2.8	125	0.00
17	Allyl chloride	3.034	3.155	-4.0	116	0.00
18 M	Methylene Chloride	1.958	1.925	1.7	105	0.00
19 M	trans-1,2-Dichloroethene	1.695	1.748	-3.1	112	0.00
20 T	Diisopropyl ether	6.550	6.836	-4.4	109	0.00
21 M	1,1-Dichloroethane	3.433	3.665	-6.8	116	0.00
22 M	cis-1,2-Dichloroethene	2.098	2.143	-2.1	111	0.00
23 M	tert-Butyl Alcohol	0.389	0.441	-13.4	114	0.00
24 M	Methyl tert-Butyl Ether	6.061	6.608	-9.0	115	0.00
25 M	Chloroform	3.523	3.929	-11.5	122	0.00
26	Cyclohexane	2.596	2.868	-10.5	162#	0.00
27 s	1,2-Dichloroethane-d4	2.493	2.799	-12.3	114	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
29	1,1-Dichloropropene	0.418	0.459	-9.8	136	0.00
30 M	2-Butanone	0.273	0.286	-4.8	108	0.00
31	2,2-Dichloropropane	0.449	0.483	-7.6	129	0.00
32 M	1,1,1-Trichloroethane	0.488	0.565	-15.8	140	0.00
33 M	Carbon Tetrachloride	0.385	0.446	-15.8	147	0.00
34 M	Benzene	1.318	1.330	-0.9	110	0.00
35	Methacrylonitrile	0.292	0.313	-7.2	114	0.01
36 M	1,2-Dichloroethane	0.513	0.579	-12.9	121	0.00
37 M	Trichloroethene	0.331	0.335	-1.2	114	0.00
38	Methylcyclohexane	0.458	0.479	-4.6	168#	0.00
39 M	1,2-Dichloropropane	0.355	0.368	-3.7	113	0.00
40	Dibromomethane	0.240	0.248	-3.3	112	0.00
41 M	Bromodichloromethane	0.434	0.468	-7.8	120	0.00
42 M	Vinyl Acetate	0.803	0.805	-0.2	110	0.00
43	Ethyl Acetate	0.522	0.536	-2.7	113	0.00
44	Isopropyl Acetate	0.850	0.862	-1.4	111	0.00
45 T	1,4-Dioxane	0.008	0.009	-12.5	104	0.00
46	Methyl methacrylate	0.411	0.421	-2.4	114	0.00
47	n-amyl Acetate	0.714	0.707	1.0	110	0.00
48 M	t-1,3-Dichloropropene	0.480	0.470	2.1	114	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
 Data File : VX036067.D
 Acq On : 07 Jun 2023 17:46
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Jun 08 07:23:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.530	0.517	2.5	108	0.00
50 M	1,1,2-Trichloroethane	0.343	0.360	-5.0	112	0.00
51	Ethyl methacrylate	0.529	0.538	-1.7	113	0.00
52	1,3-Dichloropropane	0.598	0.633	-5.9	112	0.00
53 M	Dibromochloromethane	0.315	0.338	-7.3	123	0.00
54 M	1,2-Dibromoethane	0.357	0.367	-2.8	110	0.00
55 M	2-Chloroethyl vinyl ether	0.281	0.306	-8.9	110	0.00
56 M	Bromoform	0.204	0.210	-2.9	126	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
58 M	4-Methyl-2-Pentanone	0.583	0.614	-5.3	114	0.00
59 M	2-Hexanone	0.438	0.459	-4.8	111	0.00
60 S	4-Bromofluorobenzene	0.503	0.529	-5.2	115	0.00
61 M	Tetrachloroethene	0.289	0.294	-1.7	120	0.00
62 M	Toluene	1.545	1.569	-1.6	114	0.00
63 S	Toluene-d8	1.326	1.343	-1.3	111	0.00
64 M	Chlorobenzene	0.978	0.989	-1.1	113	0.00
65	1,1,1,2-Tetrachloroethane	0.342	0.350	-2.3	118	0.00
66 M	Ethyl Benzene	1.728	1.786	-3.4	121	0.00
67 M	m/p-Xylenes	0.654	0.672	-2.8	118	0.00
68 M	o-Xylene	0.654	0.668	-2.1	114	0.00
69 M	Styrene	1.080	1.125	-4.2	118	0.00
70	Isopropylbenzene	1.670	1.774	-6.2	125	0.00
71 M	1,1,2,2-Tetrachloroethane	0.591	0.622	-5.2	114	0.00
72	1,2,3-Trichloropropane	0.515	0.530	-2.9	118	0.00
73	Bromobenzene	0.406	0.407	-0.2	113	0.00
74	n-propylbenzene	1.990	2.052	-3.1	122	0.00
75	2-Chlorotoluene	1.223	1.282	-4.8	121	0.00
76	1,3,5-Trimethylbenzene	1.418	1.500	-5.8	124	0.00
77	t-1,4-Dichloro-2-butene	0.155	0.129	16.8	108	0.00
78	4-Chlorotoluene	1.393	1.477	-6.0	122	0.00
79	tert-butylbenzene	1.369	1.435	-4.8	126	0.00
80	1,2,4-Trimethylbenzene	1.425	1.503	-5.5	122	0.00
81	sec-Butylbenzene	1.728	1.793	-3.8	130	0.00
82	p-Isopropyltoluene	1.438	1.494	-3.9	128	0.00
83 M	1,3-Dichlorobenzene	0.768	0.782	-1.8	116	0.00
84 M	1,4-Dichlorobenzene	0.771	0.791	-2.6	117	0.00
85	n-Butylbenzene	1.279	1.272	0.5	131	0.00
86 T	Hexachloroethane	0.226	0.219	3.1	134	0.00
87 M	1,2-Dichlorobenzene	0.752	0.782	-4.0	114	0.00
88	1,2-Dibromo-3-Chloropropane	0.113	0.122	-8.0	127	0.00
89	1,2,4-Trichlorobenzene	0.451	0.426	5.5	112	0.00
90	Hexachlorobutadiene	0.166	0.181	-9.0	155#	0.00
91 M	Naphthalene	1.560	1.579	-1.2	114	0.00
92	1,2,3-Trichlorobenzene	0.450	0.434	3.6	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0