



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060723\
 Data File : VX036049.D
 Acq On : 07 Jun 2023 08:24
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Tue May 23 11:36:37 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	8228	PASS
75	95	30	60	55.7	21653	PASS
95	95	100	100	100.0	38883	PASS
96	95	5	9	7.0	2711	PASS
173	174	0.00	2	0.5	128	PASS
174	95	50	100	64.0	24877	PASS
175	174	5	9	8.1	2010	PASS
176	174	95	101	97.7	24296	PASS
177	176	5	9	6.3	1530	PASS