

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060724\
 Data File : VX041784.D
 Acq On : 07 Jun 2024 12:41
 Operator : JC/MD
 Sample : P2699-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240529065-02

Quant Time: Jun 08 04:04:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	23421	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	130604	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	115644	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	58345	31.634	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.433%
60) 4-Bromofluorobenzene	11.079	95	50512	27.980	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.267%
63) Toluene-d8	8.647	98	145663	29.384	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.933%
Target Compounds						
					Qvalue	
15) Acetone	2.392	58	2654	12.484	ug/l	93
25) Chloroform	5.093	83	11697	4.379	ug/l	97
41) Bromodichloromethane	7.824	83	2114	1.109	ug/l #	98
62) Toluene	8.726	91	4562	0.793	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060724\
Data File : VX041784.D
Acq On : 07 Jun 2024 12:41
Operator : JC/MD
Sample : P2699-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
240529065-02

Quant Time: Jun 08 04:04:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 16 05:13:13 2024
Response via : Initial Calibration

