

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060724\
 Data File : VX041787.D
 Acq On : 07 Jun 2024 13:50
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 08 04:06:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	25344	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	138912	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	123060	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	61139	30.633	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.100%
60) 4-Bromofluorobenzene	11.079	95	53887	28.051	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.500%
63) Toluene-d8	8.647	98	156817	29.728	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.100%
Target Compounds						
15) Acetone	2.386	58	1267	5.508	ug/l	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060724\
Data File : VX041787.D
Acq On : 07 Jun 2024 13:50
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 08 04:06:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 16 05:13:13 2024
Response via : Initial Calibration

