

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060724\
 Data File : VX041788.D
 Acq On : 07 Jun 2024 14:14
 Operator : JC/MD
 Sample : P2777-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240605059-07-TRIP-BLANK

Quant Time: Jun 08 04:06:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	24589	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	133142	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	119034	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	59343	30.647	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.167%
60) 4-Bromofluorobenzene	11.079	95	51689	27.817	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.733%
63) Toluene-d8	8.647	98	149963	29.390	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.967%
Target Compounds						
					Qvalue	
15) Acetone	2.392	58	1037	4.646	ug/l	91
25) Chloroform	5.099	83	5197	1.853	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060724\
Data File : VX041788.D
Acq On : 07 Jun 2024 14:14
Operator : JC/MD
Sample : P2777-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
240605059-07-TRIP-BLANK

Quant Time: Jun 08 04:06:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 16 05:13:13 2024
Response via : Initial Calibration

