

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060819\
 Data File : VX010140.D
 Acq On : 08 Jun 2019 15:56
 Operator : JC/SP
 Sample : K3221-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB02-20190604

Quant Time: Jun 10 08:25:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	310935	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	466379	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	418764	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	194856	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	138146	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
35) Dibromofluoromethane	5.50	113	143945	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
50) Toluene-d8	8.71	98	546038	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
62) 4-Bromofluorobenzene	11.13	95	189006	46.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.88%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060819\
Data File : VX010140.D
Acq On : 08 Jun 2019 15:56
Operator : JC/SP
Sample : K3221-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB02-20190604

Quant Time: Jun 10 08:25:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

