

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060819\
 Data File : VX010142.D
 Acq On : 08 Jun 2019 16:43
 Operator : JC/SP
 Sample : K3257-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-20190606

Quant Time: Jun 10 08:32:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	300436	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	453106	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	408395	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	188530	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	133408	45.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.98%	
35) Dibromofluoromethane	5.50	113	141828	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
50) Toluene-d8	8.71	98	529478	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
62) 4-Bromofluorobenzene	11.13	95	183807	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
Target Compounds						
16) Acetone	2.45	43	10527	7.632	ug/l #	86
52) Toluene	8.79	92	28728	3.823	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060819\
Data File : VX010142.D
Acq On : 08 Jun 2019 16:43
Operator : JC/SP
Sample : K3257-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-20190606

Quant Time: Jun 10 08:32:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

