

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060819\
 Data File : VX010147.D
 Acq On : 08 Jun 2019 18:40
 Operator : JC/SP
 Sample : K3249-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-HN-MW24IR-20190606

Quant Time: Jun 10 09:57:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	297159	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	445129	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	405151	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	187352	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	131057	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
35) Dibromofluoromethane	5.50	113	137988	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
50) Toluene-d8	8.71	98	522804	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	11.13	95	182090	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
Target Compounds						
7) Trichlorofluoromethane	1.93	101	5074	1.338	ug/l	96
12) 1,1-Dichloroethene	2.38	96	2304	0.863	ug/l #	75
19) Methyl tert-butyl Ether	3.20	73	6945	0.810	ug/l	94
24) 1,1-Dichloroethane	3.70	63	3633	0.777	ug/l #	85
27) cis-1,2-Dichloroethene	4.60	96	1409	0.429	ug/l	72
44) Trichloroethene	7.21	130	25967	7.671	ug/l	84
64) Tetrachloroethene	9.34	164	11648	3.871	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060819\
Data File : VX010147.D
Acq On : 08 Jun 2019 18:40
Operator : JC/SP
Sample : K3249-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-HN-MW24IR-20190606

Quant Time: Jun 10 09:57:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

