

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060819\
 Data File : VX010148.D
 Acq On : 08 Jun 2019 19:03
 Operator : JC/SP
 Sample : K3249-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-DUP05-20190606

Quant Time: Jun 10 08:43:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	286077	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	430337	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	385614	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	182194	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	127083	45.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.02%	
35) Dibromofluoromethane	5.50	113	132644	49.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.94%	
50) Toluene-d8	8.71	98	499105	50.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.34%	
62) 4-Bromofluorobenzene	11.13	95	173774	46.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.54%	
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.93	101	5254	1.440	ug/l	99
12) 1,1-Dichloroethene	2.37	96	2212	0.861	ug/l #	69
19) Methyl tert-butyl Ether	3.20	73	6699	0.811	ug/l #	87
24) 1,1-Dichloroethane	3.70	63	3338	0.741	ug/l #	81
27) cis-1,2-Dichloroethene	4.61	96	1522	0.481	ug/l	66
44) Trichloroethene	7.21	130	25183	7.696	ug/l	92
64) Tetrachloroethene	9.34	164	11819	4.127	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060819\
Data File : VX010148.D
Acq On : 08 Jun 2019 19:03
Operator : JC/SP
Sample : K3249-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-DUP05-20190606

Quant Time: Jun 10 08:43:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

