

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060820\
 Data File : VX016634.D
 Acq On : 08 Jun 2020 13:35
 Operator : JC/SP
 Sample : L2894-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW004-200601

Quant Time: Jun 09 05:55:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	219540	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	349208	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	350112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	177906	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	130253	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
35) Dibromofluoromethane	5.47	113	105613	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
50) Toluene-d8	8.70	98	425598	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
62) 4-Bromofluorobenzene	11.12	95	172696	53.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.80%	
Target Compounds						
16) Acetone	2.42	43	4013	3.360	ug/l	95
27) cis-1,2-Dichloroethene	4.57	96	41296	15.277	ug/l	87
44) Trichloroethene	7.19	130	1105	0.354	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060820\
Data File : VX016634.D
Acq On : 08 Jun 2020 13:35
Operator : JC/SP
Sample : L2894-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW004-200601

Quant Time: Jun 09 05:55:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

