

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060820\
 Data File : VX016642.D
 Acq On : 08 Jun 2020 17:12
 Operator : JC/SP
 Sample : PB129340ZHE#09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129340ZHE#09

Quant Time: Jun 09 06:18:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	234273	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	368965	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	379781	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	190062	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	135019	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
35) Dibromofluoromethane	5.47	113	112428	49.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.98%	
50) Toluene-d8	8.70	98	458736	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
62) 4-Bromofluorobenzene	11.12	95	182314	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
Target Compounds						
16) Acetone	2.42	43	19880	15.596	ug/l	96
20) Methylene Chloride	2.84	84	6070	2.302	ug/l	91
43) Isopropyl Acetate	6.42	43	65842	10.064	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060820\
Data File : VX016642.D
Acq On : 08 Jun 2020 17:12
Operator : JC/SP
Sample : PB129340ZHE#09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129340ZHE#09

Quant Time: Jun 09 06:18:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

