

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060820\
 Data File : VX016644.D
 Acq On : 08 Jun 2020 17:57
 Operator : JC/SP
 Sample : PB129340ZHE#11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129340ZHE#11

Quant Time: Jun 09 06:21:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	221826	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	359953	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	364316	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	184110	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	132223	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
35) Dibromofluoromethane	5.46	113	109225	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
50) Toluene-d8	8.70	98	447173	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
62) 4-Bromofluorobenzene	11.12	95	178276	53.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.96%	
Target Compounds						
16) Acetone	2.42	43	19330	16.016	ug/l	98
18) Methyl Acetate	2.75	43	3165	1.022	ug/l	100
20) Methylene Chloride	2.84	84	6729	2.695	ug/l	86
43) Isopropyl Acetate	6.41	43	67964	10.649	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060820\
Data File : VX016644.D
Acq On : 08 Jun 2020 17:57
Operator : JC/SP
Sample : PB129340ZHE#11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129340ZHE#11

Quant Time: Jun 09 06:21:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

