

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060820\
 Data File : VX016646.D
 Acq On : 08 Jun 2020 18:42
 Operator : JC/SP
 Sample : L2949-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jun 09 06:25:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	219228	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	350501	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	358635	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	181704	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	131218	48.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.40%	
35) Dibromofluoromethane	5.47	113	106331	49.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.54%	
50) Toluene-d8	8.70	98	435482	51.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.68%	
62) 4-Bromofluorobenzene	11.12	95	175345	54.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.04%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060820\
Data File : VX016646.D
Acq On : 08 Jun 2020 18:42
Operator : JC/SP
Sample : L2949-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jun 09 06:25:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

