


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX060820\
 Data File : VX016626.D
 Acq On : 08 Jun 2020 09:15
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Title : SW846 8260
 Last Update : Wed May 27 16:31:05 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.6	8612	PASS
75	95	30	60	48.0	26416	PASS
95	95	100	100	100.0	55029	PASS
96	95	5	9	6.9	3824	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	87.4	48080	PASS
175	174	5	9	7.6	3640	PASS
176	174	95	101	98.1	47168	PASS
177	176	5	9	6.3	2967	PASS