

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022555.D
 Acq On : 08 Jun 2021 11:46
 Operator : JC/MD
 Sample : VX0608WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0608WBL01

Quant Time: Jun 09 03:58:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60191	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	121023	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	109109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42586	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50335	47.040	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.080%
35) Dibromofluoromethane	5.391	113	36571	48.298	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.600%
50) Toluene-d8	8.653	98	148829	49.260	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.520%
62) 4-Bromofluorobenzene	11.086	95	54040	48.121	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022555.D
Acq On : 08 Jun 2021 11:46
Operator : JC/MD
Sample : VX0608WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0608WBL01

Quant Time: Jun 09 03:58:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

