

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022559.D
 Acq On : 08 Jun 2021 13:25
 Operator : JC/MD
 Sample : PB136857TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136857TB

Quant Time: Jun 09 03:59:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59446	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	121797	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	108474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42432	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50061	47.371	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.740%
35) Dibromofluoromethane	5.397	113	36979	48.526	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.060%
50) Toluene-d8	8.653	98	150986	49.657	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.320%
62) 4-Bromofluorobenzene	11.085	95	51813	45.845	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.700%
Target Compounds						
						Qvalue
16) Acetone	2.385	43	9070	19.100	ug/l	# 86
20) Methylene Chloride	2.794	84	3798	3.831	ug/l	# 78
43) Isopropyl Acetate	6.348	43	11361	4.634	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022559.D
Acq On : 08 Jun 2021 13:25
Operator : JC/MD
Sample : PB136857TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136857TB

Quant Time: Jun 09 03:59:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

