

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022561.D
 Acq On : 08 Jun 2021 14:12
 Operator : JC/MD
 Sample : PB136878ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136878ZHE#01

Quant Time: Jun 09 03:59:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	61777	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	126208	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	117356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51981	47.331	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.660%
35) Dibromofluoromethane	5.397	113	37473	47.456	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.920%
50) Toluene-d8	8.653	98	158164	50.200	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.400%
62) 4-Bromofluorobenzene	11.085	95	56959	48.637	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.280%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	8488	17.200	ug/l	91
20) Methylene Chloride	2.800	84	4127	4.006	ug/l	88
43) Isopropyl Acetate	6.355	43	11526	4.537	ug/l	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022561.D
Acq On : 08 Jun 2021 14:12
Operator : JC/MD
Sample : PB136878ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136878ZHE#01

Quant Time: Jun 09 03:59:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

