

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022564.D
 Acq On : 08 Jun 2021 15:20
 Operator : JC/MD
 Sample : PB136878ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136878ZHE#04

Quant Time: Jun 09 04:00:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	58421	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	122503	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	111894	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43804	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	50894	49.004	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.000%		
35) Dibromofluoromethane	5.404	113	36580	47.726	ug/l	0.01
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.460%		
50) Toluene-d8	8.653	98	152785	49.959	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.920%		
62) 4-Bromofluorobenzene	11.085	95	52881	46.520	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.040%		
Target Compounds						Qvalue
16) Acetone	2.386	43	8446	18.098	ug/l	98
18) Methyl Acetate	2.715	43	1436	1.116	ug/l #	75
20) Methylene Chloride	2.788	84	4412	4.528	ug/l #	89
43) Isopropyl Acetate	6.348	43	11569	4.692	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022564.D
Acq On : 08 Jun 2021 15:20
Operator : JC/MD
Sample : PB136878ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136878ZHE#04

Quant Time: Jun 09 04:00:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

