

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022565.D
 Acq On : 08 Jun 2021 15:44
 Operator : JC/MD
 Sample : PB136878ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136878ZHE#05

Quant Time: Jun 09 04:00:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	62249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	127447	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52590	47.523	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.040%
35) Dibromofluoromethane	5.397	113	37373	46.869	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.740%
50) Toluene-d8	8.653	98	159198	50.037	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.080%
62) 4-Bromofluorobenzene	11.085	95	57367	48.509	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.020%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	8246	16.583	ug/l #	82
18) Methyl Acetate	2.715	43	1708	1.246	ug/l	95
20) Methylene Chloride	2.794	84	3396	3.271	ug/l	94
43) Isopropyl Acetate	6.348	43	10107	3.940	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022565.D
Acq On : 08 Jun 2021 15:44
Operator : JC/MD
Sample : PB136878ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136878ZHE#05

Quant Time: Jun 09 04:00:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

