

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022569.D
 Acq On : 08 Jun 2021 17:18
 Operator : JC/MD
 Sample : PB136878ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136878ZHE#09

Quant Time: Jun 09 04:01:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59089	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	120401	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43486	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49011	46.657	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.320%
35) Dibromofluoromethane	5.397	113	35183	46.705	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.400%
50) Toluene-d8	8.653	98	148670	49.462	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.920%
62) 4-Bromofluorobenzene	11.085	95	52440	46.938	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.880%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	7220	15.296	ug/l	94
18) Methyl Acetate	2.715	43	1565	1.202	ug/l #	79
20) Methylene Chloride	2.794	84	1994	2.023	ug/l #	84
43) Isopropyl Acetate	6.348	43	11958	4.934	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022569.D
Acq On : 08 Jun 2021 17:18
Operator : JC/MD
Sample : PB136878ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136878ZHE#09

Quant Time: Jun 09 04:01:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

