

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022571.D
 Acq On : 08 Jun 2021 18:05
 Operator : JC/MD
 Sample : PB136878ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136878ZHE#11

Quant Time: Jun 09 04:01:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59703	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	122065	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	108592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51429	48.456	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.920%
35) Dibromofluoromethane	5.397	113	36575	47.891	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.780%
50) Toluene-d8	8.653	98	153648	50.421	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.840%
62) 4-Bromofluorobenzene	11.085	95	52027	45.933	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.860%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	7536	15.801	ug/l	99
18) Methyl Acetate	2.715	43	1434	1.090	ug/l	# 80
20) Methylene Chloride	2.800	84	2013	2.022	ug/l	# 84
43) Isopropyl Acetate	6.355	43	11811	4.807	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022571.D
Acq On : 08 Jun 2021 18:05
Operator : JC/MD
Sample : PB136878ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136878ZHE#11

Quant Time: Jun 09 04:01:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

