

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022575.D
 Acq On : 08 Jun 2021 19:39
 Operator : JC/MD
 Sample : M2625-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B12(138-142)

Quant Time: Jun 09 04:01:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	60987	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	123464	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46057	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52589	48.505	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.020%
35) Dibromofluoromethane	5.397	113	37482	48.522	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.040%
50) Toluene-d8	8.653	98	160676	52.130	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.260%
62) 4-Bromofluorobenzene	11.085	95	56799	49.578	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.160%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6083	12.486	ug/l	93
18) Methyl Acetate	2.715	43	2188	1.629	ug/l #	80
20) Methylene Chloride	2.794	84	4025	3.957	ug/l #	91
43) Isopropyl Acetate	6.355	43	11340	4.563	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022575.D
Acq On : 08 Jun 2021 19:39
Operator : JC/MD
Sample : M2625-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B12(138-142)

Quant Time: Jun 09 04:01:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

