

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060821\  
 Data File : VX022576.D  
 Acq On : 08 Jun 2021 20:02  
 Operator : JC/MD  
 Sample : M2625-22  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 18-B12(142-146)

Quant Time: Jun 09 04:02:05 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060721W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jun 07 14:27:50 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.562	168	57310	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	121002	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	108882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	50657	49.721	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.440%
35) Dibromofluoromethane	5.397	113	35375	46.726	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.460%
50) Toluene-d8	8.653	98	148125	49.036	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.080%
62) 4-Bromofluorobenzene	11.085	95	52489	46.748	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.500%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	9898	21.621	ug/l	97
18) Methyl Acetate	2.715	43	1511	1.197	ug/l	# 83
20) Methylene Chloride	2.794	84	4883	5.109	ug/l	# 84
43) Isopropyl Acetate	6.354	43	9962	4.090	ug/l	93
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX060821\  
Data File : VX022576.D  
Acq On : 08 Jun 2021 20:02  
Operator : JC/MD  
Sample : M2625-22  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
18-B12(142-146)

Quant Time: Jun 09 04:02:05 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060721W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jun 07 14:27:50 2021  
Response via : Initial Calibration

