

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022577.D
 Acq On : 08 Jun 2021 20:26
 Operator : JC/MD
 Sample : M2625-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B12(146-150)

Quant Time: Jun 09 04:02:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	58525	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	118339	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	108362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41754	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49597	47.670	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.340%
35) Dibromofluoromethane	5.397	113	35923	48.518	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.040%
50) Toluene-d8	8.653	98	147967	50.086	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.180%
62) 4-Bromofluorobenzene	11.085	95	51152	46.583	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.160%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	10644	22.768	ug/l	92
18) Methyl Acetate	2.715	43	1482	1.149	ug/l	# 72
20) Methylene Chloride	2.794	84	4429	4.537	ug/l	# 85
43) Isopropyl Acetate	6.354	43	9834	4.129	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022577.D
Acq On : 08 Jun 2021 20:26
Operator : JC/MD
Sample : M2625-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B12(146-150)

Quant Time: Jun 09 04:02:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

