

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022578.D
 Acq On : 08 Jun 2021 20:49
 Operator : JC/MD
 Sample : M2625-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B12(150-154)

Quant Time: Jun 09 04:02:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56213	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	117221	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	106666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41051	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49245	49.278	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.560%
35) Dibromofluoromethane	5.391	113	35337	48.182	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.360%
50) Toluene-d8	8.653	98	145710	49.792	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.580%
62) 4-Bromofluorobenzene	11.085	95	50160	46.115	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.220%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	6580	14.653	ug/l	97
18) Methyl Acetate	2.715	43	1608	1.298	ug/l	# 89
20) Methylene Chloride	2.794	84	4762	5.079	ug/l	# 85
43) Isopropyl Acetate	6.355	43	11204	4.749	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022578.D
Acq On : 08 Jun 2021 20:49
Operator : JC/MD
Sample : M2625-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B12(150-154)

Quant Time: Jun 09 04:02:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

