

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022579.D
 Acq On : 08 Jun 2021 21:13
 Operator : JC/MD
 Sample : M2609-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-AQ-FT1-060321

Quant Time: Jun 09 04:02:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56805	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	118971	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106751	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41454	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50897	50.401	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.800%	
35) Dibromofluoromethane	5.403	113	35459	47.637	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.280%	
50) Toluene-d8	8.653	98	146411	49.296	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.600%	
62) 4-Bromofluorobenzene	11.085	95	51258	46.431	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.860%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	15720	34.643	ug/l	95
18) Methyl Acetate	2.715	43	1181	0.944	ug/l	# 83
25) 2-Butanone	4.574	43	5154	6.873	ug/l	# 89
27) cis-1,2-Dichloroethene	4.495	96	885	0.929	ug/l	86
30) Chloroform	5.111	83	987	0.652	ug/l	# 78
44) Trichloroethene	7.129	130	1847	2.204	ug/l	93
52) Toluene	8.720	92	2557	1.146	ug/l	92
68) m/p-Xylenes	10.305	106	1183	0.797	ug/l	# 65
69) o-Xylene	10.646	106	901	0.621	ug/l	82
80) 1,3,5-Trimethylbenzene	11.451	105	1047	0.363	ug/l	83
84) 1,2,4-Trimethylbenzene	11.756	105	2129	0.733	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022579.D
Acq On : 08 Jun 2021 21:13
Operator : JC/MD
Sample : M2609-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-AQ-FT1-060321

Quant Time: Jun 09 04:02:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

