

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022588.D
 Acq On : 09 Jun 2021 01:31
 Operator : JC/MD
 Sample : M2625-27
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B12(162-166)

Quant Time: Jun 09 04:04:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59411	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	120956	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	111811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51531	48.790	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.580%
35) Dibromofluoromethane	5.397	113	36064	47.655	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.300%
50) Toluene-d8	8.653	98	152152	50.388	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.780%
62) 4-Bromofluorobenzene	11.085	95	53756	47.895	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.780%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	5652	11.909	ug/l	99
20) Methylene Chloride	2.794	84	4204	4.243	ug/l	# 88
43) Isopropyl Acetate	6.354	43	11344	4.659	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022588.D
Acq On : 09 Jun 2021 01:31
Operator : JC/MD
Sample : M2625-27
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B12(162-166)

Quant Time: Jun 09 04:04:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

