

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022589.D
 Acq On : 09 Jun 2021 01:55
 Operator : JC/MD
 Sample : M2625-28
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B12(166-170)

Quant Time: Jun 09 04:04:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60956	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	124061	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51834	47.833	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.660%
35) Dibromofluoromethane	5.397	113	36800	47.410	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.820%
50) Toluene-d8	8.653	98	156662	50.583	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.160%
62) 4-Bromofluorobenzene	11.085	95	56688	49.243	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.480%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	9303	19.106	ug/l	97
20) Methylene Chloride	2.794	84	3882	3.818	ug/l #	85
43) Isopropyl Acetate	6.348	43	9097	3.643	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022589.D
Acq On : 09 Jun 2021 01:55
Operator : JC/MD
Sample : M2625-28
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B12(166-170)

Quant Time: Jun 09 04:04:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

