

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
 Data File : VX022593.D
 Acq On : 09 Jun 2021 03:29
 Operator : JC/MD
 Sample : M2625-32
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 18-B12(182-186)

Quant Time: Jun 09 04:05:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 07 14:27:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56463	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	114490	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	106058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	48417	48.235	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.480%
35) Dibromofluoromethane	5.403	113	34934	48.768	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.540%
50) Toluene-d8	8.653	98	145143	50.782	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.560%
62) 4-Bromofluorobenzene	11.085	95	50175	47.229	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.460%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	6426	14.247	ug/l	# 82
18) Methyl Acetate	2.721	43	1340	1.077	ug/l	# 69
20) Methylene Chloride	2.794	84	4344	4.613	ug/l	# 88
43) Isopropyl Acetate	6.354	43	11733	5.091	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060821\
Data File : VX022593.D
Acq On : 09 Jun 2021 03:29
Operator : JC/MD
Sample : M2625-32
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
18-B12(182-186)

Quant Time: Jun 09 04:05:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060721W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 07 14:27:50 2021
Response via : Initial Calibration

